

CIED guideline recommendations in the setting of advisories

Michael Sawyer¹ and Paul Gould¹

¹Princess Alexandra Hospital

December 16, 2021

Abstract

Subcutaneous implantable cardioverter defibrillators (S-ICDs) are currently indicated for patients who meet ICD implantation criteria for the prevention of sudden cardiac death (SCD). In December 2020, a unique cardiac implantable electronic device (CIED) situation arose as both the S-ICD generator and electrode were under a device advisory. We would recommend all future CIED implantation guidelines receive a caveat regarding checking current CIED advisories, in order to avoid future similar scenarios.

Hosted file

SICD letter.docx available at <https://authorea.com/users/451510/articles/549712-cied-guideline-recommendations-in-the-setting-of-advisories>